

Questions and Answers

1.1. Question:

- Diesel generators- it weights 600kg and its 1,7m long it can't go under the contain/truck should the bid include the cost for a trailer?

Answer: The specifications were based on the truck therefore the bidder should apply the necessary requirements for the containers and include the cost of a trailer in their proposal.

- Chest X-Ray stand- is a motorized wall stand required?

Answer: Provide a quotation for both the motorized and non-motorized chest X-Ray stand

- Solar Panels- is there a need for both the solar batteries and the UPS(electricity)?

Answer: Solar panels are a high risk for being stolen, it may be better to use electricity.

1.2. Question:

- What is the delivery address for the documents and for the containers?

Answer: If lockdown continues the documents should be submitted electronically however the hard copy should be kept for when needed.

Answer: The delivery should be costed according to the districts that are listed in the tender documents – district offices should be used to estimate delivery costs.

1.3. Question:

- Only the specification of the inside of the containers are clearly prescribed there are no structural specifications that is no drawings provided.
- What is ISO 6 container?
- What kind of insulation is needed because there are 30 kinds of insulation available?

Answer: the bidder must provide the drawings we only give specifications for what we need the container to include.

Answer; ISO 6 - international nomenclature with specifications on transport container s

Answer: The bidder must produce a container that is fitted with medical equipment and they must conform to the provided specifications.

1.4. Question:

- What are the required timelines for the delivery of the containers?

Answer: Bidder to state in the bid document the timelines for delivery of equipped containers.

- Medical equipment should it be sourced internationally or locally?

Answer: The equipment can be sourced locally or internationally as long as they are SABS approved

1.5. Question:

- Can the delivery be divided up into smaller batches i.e. 5 units at a time or must we deliver all 15 at the same time?
- Answer: The delivery can be in batches and the bidder must provide the implementation/delivery plan to enable tracking of the implementation. This will assist with cash flow availability for the payment since payment will be after delivery.

1.6. Question:

- The AED must it be erect bucky or table bucky?

Answer: The container must have both the erect and the table or supine ability to perform x-rays.

- Will there be extension if lockdown continues?

Answer: There will be no extension for the submission of the proposal the submission date stands.

1.7. Question:

- Are Joint Ventures allowed?

Answer: Joint ventures are allowed only if they follow the guidance stated in the PFMA and all the legal documents are provided.

1.8. Question:

- Warranty period? Service and maintenance agreement how many years?

Answer: The bidder should state the warranty and the service and maintenance plan for the equipment

1.9. Question:

- For how long should the bid be valid for? ZAR/\$ exchange will it be considered due to the volatile of rand.

Answer: The bid will be valid for 90 days and we are working with the ZAR.